



EU - TYPE EXAMINATION CERTIFICATE



No. SK 13 - 023 MI-004 Rev. 4

This revision replaces all previous versions of this Certificate in full wording

Issued by **Slovenská legálna metrologia, n. o.** Notified Body number **1432**
Hviezdoslavova 31
974 01 Banská Bystrica
Slovak Republic

In accordance with Annex II, Module B to Government Ordinance of the Slovak Republic No. 145/2016 Coll. relating to the making available on the market of measuring instruments, which implements, in Slovakia, the Directive 2014/32/EU of the European Parliament and of the Council of 26 February 2014 on the harmonisation of the laws of the Member States relating to the making available on the market of measuring instruments as later amended (MID).

Applicable essential requirements Annex I and Annex VI to MID

Manufacturer **Atlas Sayac ve Olcu Aletleri Yazilim Otm. San. A.S.**
Mimarsinan OSB 23. Cad. No: 70
Melikgazi, Kayseri /TURKEY

Applicant **Manufacturer**

Measuring instrument **Thermal energy meter**

Type **UKM... or VHU... or DAF... or ULTRA...**
Trade mark see Descriptive annex
SW version 1.66

Environment classes
- climatic (+5 to +55) °C
- mechanical M1
- electromagnetic E1


Description and documentation The principal technical and metrological data, characteristics, instrument description and approval conditions are set out in the Descriptive annex to this EU - type examination certificate (20 pages), which is part of this EU - type examination certificate. The test reports, designs, schematic diagrams and documentation used during certification process are recorded under reference folder Atlas_UKM_00 to UKM_04.

Valid until **1 October 2023**

Date of issue **23 January 2018**



Ing. Štefan Král, PhD.
Representative of Notified Body



Where the instrument is subject to other Directives covering other aspects, this EU - type examination certificate is valid, assuming that the instrument conforms to the provisions of those Directives. Without written permission of the notified body this certificate may be reproduced only as a whole.

