



EU - TYPE EXAMINATION CERTIFICATE



No. SK 14 - 092 MI-001 Rev. 3

This revision replaces all previous versions of this Certificate in full wording

Issued by	Slovenská legálna metrologia, n. o. Hviezdoslavova 31 974 01 Banská Bystrica Slovak Republic	Notified Body number 1432
In accordance with	Annex II, Module B to Regulation of the Government of the Slovak Republic No 145/2016 Coll. on making available of measuring instruments on the market, in the wording of the Regulation of the Government of the Slovak Republic No 328/2019 Coll., which implements, in Slovakia, the Directive 2014/32/EU of the European Parliament and of the Council of 26 February 2014 on the harmonisation of the laws of the Member States relating to the making available on the market of measuring instruments as later amended (MID).	
Applicable essential requirements	Annex I and Annex III to MID	
Manufacturer	Ningbo Aimei Meter Manufacture Co., Ltd. 68, West Town Road, Shangtian Town, Fenghua Zhejiang, China 315511	
Applicant	Manufacturer	
Measuring instrument	Water meter	
	Type	MD-A, MD-AP
	Trade mark	see Descriptive annex
	Environment classes	
	- climatic	(-10 to +55)°C
	- mechanical	M1
	- electromagnetic	E1
Description and documentation	The principal technical and metrological data, characteristics, instrument description and approval conditions are set out in the Descriptive annex to this EU - type examination certificate (14 pages), which is part of this EU - type examination certificate. The test reports, designs, schematic diagrams and documentation used during certification process are stored in the reference folder Ningbo Aimei_MD-A_AP_00 to 03.	

Valid until **5 March 2034**

Date of issue **5 March 2024**



Ing. Dušan Šmigura, PhD.
Representative of Notified Body



Where the instrument is subject to other Directives covering other aspects, this EU - type examination certificate is valid, assuming that the instrument conforms to the provisions of those Directives. Without written permission of the notified body this certificate may be reproduced only as a whole.